

PRODUCT DATA SHEET

Standing: 2024-11-04

LAUDA Hydro H 2 P

Tissue float bath 115 V; 60 Hz

Part Number: L003077

Features

- Tissue float bath for stretching and drying cut tissue samples
- Temperature indication by control thermometer at the bath edge
- Temperature adjustment via rotary knob with temperature scale
- 1-point temperature controller
- Water deficiency protection
- Aluminium bath interior, black anodised
- Control thermometer with holder as standard



Working temperature min.
25 °C



Working temperature max.
80 °C

Technical Features

Working temperature range	25 ... 80 °C
Ambient temperature range	10 ... 40 °C
Temperature stability	0.5 ± K
Heater power max.	0.3 kW
Power consumption max.	0.3 kW
Current max.	16 A
Size of bath (ØxH)	200 x 60 mm
Overall dimensions (WxDxH)	280 x 280 x 100 mm
Weight	2 kg
Power supply	115 V; 60 Hz
Power plug	Power cord with plug (NEMA 5-15P)

LAUDA DR. R. WOBSEY GMBH & CO. KG
Laudaplatz 1 • 97922 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0
info@lauda.de • www.lauda.de
WEEE-Reg.-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:
LAUDA DR. R. WOBSEY Verwaltungs-GmbH
Sitz Lauda-Königshofen
Registergericht Mannheim • HRB 560226

Geschäftsführer:
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,
Dr. Marc Stricker
Beirat: Dr. Gerhard Wobser